

October 14, 2024

Mr. Niederhaus Waldemar
ECOM Instruments GmbH
Industriestraße 2
97959 Assamstadt
Germany

Subject: Fluke 28 II EX
Listing Number E115956; Eurofins E&E NA Project Number 132786; CML PRJ-17592
Safety Standards: • UL 60079-0 (Ed 7) 2019
UL 60079-11 (Ed 6) 2013
CSA 60079-0 (Ed 4) 2019
CSA 60079-11 (Ed 2) 2014
UL / CSA 61010-1 (Ed 3) 2012
UL / CSA 61010-2-033 (Ed 1) 2014

Dear Mr. Waldemar:

Congratulations on successfully completing the MET Mark Certification process for the Fluke 28 II EX. ECOM Instruments GmbH may begin to apply the MET Mark on the previously identified product at this time in accordance with the MET Mark Applicant Contract. The report covering the above stated product is forthcoming.

Thank you for the opportunity to perform this service for ECOM Instruments GmbH. We look forward to future opportunities with your company.

Sincerely,

Eurofins Electrical & Electronic Testing NA, Inc.



Jeremy Maxwell
Certification Officer,
Safety Laboratory



Canadian Certification has been granted under a System 3 program as defined in ISO/IEC 17067.

NRTL

Equipment Specifications					
Hazardous Location Marking Code	: U.S. and Canada Class I Division 1 Group A, B, C, D T4				
	<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 50%; text-align: center;">U.S.</td> <td style="width: 50%; text-align: center;">Canada</td> </tr> <tr> <td style="text-align: center;">Class I Zone 1 AEx ia IIC T4 Gb</td> <td style="text-align: center;">Ex ia IIC T4 Gb</td> </tr> </table>	U.S.	Canada	Class I Zone 1 AEx ia IIC T4 Gb	Ex ia IIC T4 Gb
U.S.	Canada				
Class I Zone 1 AEx ia IIC T4 Gb	Ex ia IIC T4 Gb				
Entity Parameters	: See General Product Information section below				
Rating	: Battery powered				
Installation and use classification	: Hand-held				
Ingress protection per IEC 60529	: IP6X				
Ingress protection per 60079-0 (preconditioning)	: N/A				
Ingress protection per UL 50	: N/A				
Rated ambient temperature range (°C)	: Depends on battery, see General Product Information below				