

ECOM Instruments GmbH  
 Industriestraße 2  
 97959 Assamstadt  
 Germany  
 Phone +49 6294 4224-0  
 Fax +49 6294 4224-100

No. / Nr.: DOC-5313A  
 Date / Datum: 2021-11-24

Copyright ECOM Instruments GmbH  
 www.ecom-ex.com

### Declaration of conformity / Konformitätserklärung

We, ECOM Instruments GmbH declare under our sole responsibility that the **products** listed below are in conformity with the listed **European Directives** and **standards**.

Die ECOM Instruments GmbH erklärt hiermit in alleiniger Verantwortung, dass die unten gelisteten **Produkte** den genannten **Europäischen Richtlinien und Normen** entsprechen.

### Products / Produkte

Product / Produkt	Item number	Description / Beschreibung
Fluke 568 EX ATEX / IECEX	70141507	Infrared Thermometer
FLUKE 568 EX ATEX / IECEX	70123725	Infrared Thermometer

### Directives and Standards / Richtlinien und Normen

EU-Directive EU-Richtlinie	Standards Normen
ATEX 2014/34/EU (L96/309-356)	EN 60079-11:2012-01 EN 60079-28:2015-09 EN IEC 60079-0:2018-07
EMC 2014/30/EU (L96/79-106)	EN 61326-1:2013-01 EN 61326-2-2:2013-01
RoHS 2011/65/EU (L174/88-110)	EN IEC 63000:2018-12

### Affixed CE Marking / Angebrachte CE-Kennzeichnung



### Signatures / Unterschriften

Assamstadt, 2021-11-24

Roof Wessels  
 CEO ECOM Instruments GmbH

ppa. Jörg Hartleb  
 COO ECOM Instruments GmbH

### ANNEX ATEX

**Notified Body QM-System** / Notifizierte Stelle des QM-Systems  
 Physikalisch Technische Bundesanstalt (0102)  
 Bundesallee 100  
 38116 Braunschweig  
 Germany

### Marking and Certificates / Kennzeichnung und Zertifikate

Marking Kennzeichnung	Certificate Zertifikat	Issuer ID Aussteller ID
Ex II 2 G	EPS 13 ATEX 1 525 X	2004

### Key for Issuer ID / Schlüssel zur Aussteller ID

ID	Issuer / Aussteller
2004	Bureau Veritas Consumer Products Services Germany GmbH Wilhelm-Hennemann-Straße 8 19061 Schwerin Germany