

ECOM Instruments GmbH
 Industriestraße 2
 97959 Assamstadt
 Germany
 Phone +49 6294 4224-0
 Fax +49 6294 4224-100

No: DOC-5377A
 Date: 2021-11-24

Copyright ECOM Instruments GmbH
www.ecom-ex.com

■ Declaration of conformity

We, ECOM Instruments GmbH declare under our sole responsibility that the **products** listed below are in conformity with the listed **UK Regulations** as indicated below and amended by **UK SI 2019 No. 696**, and **standards**.

■ Products

| Product | Item number | Description |
|---------------------------|-------------|----------------------|
| Fluke 568 EX ATEX / IECEX | 70141507 | Infrared Thermometer |
| FLUKE 568 EX ATEX / IECEX | 70123725 | Infrared Thermometer |

■ Regulations and Standards

| UK Regulation | Standards |
|----------------------------|--|
| UK SI 2012 No. 3032 (RoHS) | EN IEC 63000:2018-12 |
| UK SI 2016 No. 1091 (EMC) | EN 61326-1:2013-01 EN 61326-2-2:2013-01 |
| UK SI 2016 No. 1107 (EX) | EN 60079-11:2012-01 EN 60079-28:2015-09 EN IEC 60079-0:2018-07 |

■ Affixed UKCA Marking



■ Signatures

Assamstadt, 2021-11-24

Rooft Wessels
 CEO ECOM Instruments GmbH

ppa. Jörg Hartleb
 COO ECOM Instruments GmbH

■ ANNEX UK SI 2016 No. 1107 (EX)

Approved Body QM-System:

EUROFINS E&E CML LIMITED (2503)
 Newport Business Park - New Port Road
 CH65 4LZ Ellesmere Port
 United Kingdom

Marking and Certificates

| Marking | Certificate | ID |
|-----------|------------------|------|
| EX II 2 G | CML 21UKEX21002X | 2503 |

Key for ID

| ID | Responsible Body |
|------|--|
| 2503 | EUROFINS E&E CML LIMITED Newport Business Park - New Port Road CH65 4LZ Ellesmere Port United Kingdom |